CHECKLIST

Check source storage checklist

- [ ] Place the source in a metal container with a secure lid
- [ ] Line the container with foam so the source sits in the centre
- [ ] Label the outside with "RADIOACTIVE" and your contact details
- [ ] Store the container at least one metre from occupied areas
- [ ] Measure the dose rate at one metre to confirm it stays near background
- [ ] Log the date and reason each time you open the container
- [ ] Keep sources separate from detectors and other equipment
- [ ] Inspect the container and label every six months
- [ ] Inform others who share the space where the source is stored
- [ ] Arrange proper disposal when the source is no longer needed